

From: [REDACTED]
To: [REDACTED] -121 BMG; Müller, Thomas –AL 1 BMG
Cc: [REDACTED] -UAL 31 BMG; COVID-19-vaccines; [REDACTED] - PG Impfstoffe COVID-19 BMG;
Subject: AW: Eilt Silikonfrei Spritzen Biontech
Date: Donnerstag, 31. Dezember 2020 17:57:50
Attachments: [image001.png](#)
[image002.jpg](#)
[BN.png](#)
[LI-In-Bug.png](#)
[Twitter_Social_Icon_Rounded_Square_Color.png](#)

Lieber Herr Müller, [REDACTED]

BioNTech macht keine Angaben, was die Art der Spritzen betrifft. Sie können jedweden Sprizentypus verwenden. Die Entscheidung ist damit eine rein klinische Entscheidung.

Die Nadel sollte 21G oder kleiner sein. 0.3ml sollten entsprechend genau aufzuziehen sein.

Besten Dank und Grüße,

[REDACTED]

PS: Die unten angesprochen Aussagen werden wir entsprechend abklären.

[REDACTED]



Anschrift:
BioNTech SE
An der Goldgrube 12
55131 Mainz
Germany



Sitz der Gesellschaft: BioNTech SE, An der Goldgrube 12, 55131 Mainz, Germany
Amtsgericht Mainz HRB 48720 • USt.-IdNr. DE 263 382 495 • EoriNr: DE 1070347
Vorstand: Prof. Dr. Ugur Sahin • Sean Marett • Dr. Sierk Poetting • Ryan Richardson • PD Dr. Özlem Türeci

The information contained in this email is intended for the personal and confidential use of the addressee only. It may also contain privileged information. If you are not the intended recipient then you are hereby notified that you have received this document in error and that any review, distribution or copying of this document is strictly prohibited. If you have received this communication in error, please notify the sender.

Please protect our environment and consider whether you need to print this e-mail.

Von: [REDACTED] -121 BMG [mailto:[REDACTED]@bmg.bund.de]
Gesendet: Donnerstag, 31. Dezember 2020 12:49
An: [REDACTED]@biontech.de>
Cc: [REDACTED] -UAL 31 BMG <[REDACTED]@bmg.bund.de>; COVID-19-vaccines <COVID-19-vaccines@bmg.bund.de>; [REDACTED] - PG Impfstoffe COVID-19 BMG <[REDACTED]@bmg.bund.de>
Betreff: WG: Eilt Silikonfrei Spritzen Biontech
Priorität: Hoch

Attention: This is an external email. Please verify sender before replying, clicking links or opening any attachments.

Liebes BioNTech-Team,

nach unserem Stand sind bei Comirnaty nicht zwingend silikonölfreie Spritzen zu verwenden – diese Frage hatten wir ja eigentlich seit einiger Zeit bereits geklärt. Offenbar war so aber die Aussage bzw. zumindest Verständnis bei einigen Teilnehmern des heutigen Webinars, siehe unten. Könnten Sie das bitte kurzfristig klarstellen? Wir gehen weiter davon aus, dass Silikonfreiheit nicht notwendig ist, richtig?

Danke und freundliche Grüße

[REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]